AMETEK®
ULTRA PRECISION TECHNOLOGIES

Solartron Metrology

SI7500

Loading and Saving Settings and Parts

> Plug a USB flash memory stick into the SI7500's USB port. If loading an existing Settings or Part file onto the SI7500, ensure it is mounted onto the memory stick's root directory.

> Turn on the unit and enter the setup menu.

> Select the Supervisor screen and enter the supervisor password.

> Press Finish twice to exit the Setup menu.

> Re-enter the Supervisor screen in the Setup menu.

> Select the Supervisor password field and press the softkey below the "Load.." or "Save.." option.

> Select the desired data type to save or load by pressing the softkey below the displayed options.

> Follow the on-screen prompts to finalise the data transfer.

> Done! Press "OK" to exit to the setup menu.